



中国认可
国际互认
校准
CALIBRATION
CNAS L9002



235A007020008



安正计量检测有限公司

ANZHENG METROLOGY AND TEST CO., LTD.

校准证书

Calibration Certificate

证书编号:

EE20233694026

Certificate No.

委托单位

湖北仪天成电力设备有限公司

Client

委托地址

武汉市东湖高新区光谷大道303号光谷·芯中心1-05栋文凤楼1-4层01

Address

器具名称

带电电缆识别仪

Description

制造厂家

湖北仪天成电力设备有限公司

Manufactory

型号/规格

YTC601A

Type or size

器具编号

231506

Number of sample

校准结果

所校准项目合格

Calibration Conclusion



批准人:

李明祥

Authorized by

核验员:

张海明

Checked by

校准员:

李丹

Calibrated by

接收日期 2023 年 05 月 28 日
Date of Receipt Year Month Day

校准日期 2023 年 05 月 30 日
Date of Calibration Year Month Day

发布日期 2023 年 06 月 01 日
Date of Publication Year Month Day

地址: 福建省福州市仓山区仓山科技园1区02号1#楼101室

ADD: Room 101, Building 1#, No 02, District 1, cangshan Science and Technology Park, cangshan district, Fuzhou city, Fujian province

邮编(Post Code): 350026

传真(Fax): 0591-83591276

服务电话(Tel): 0591-88030652

网址(Web): www.fjazjl.com



校准说明

DIRECTIONS OF CALIBRATION

1、本实验室出具的数据均可溯源至国家计量基(标)准和国际单位制(SI)。

All data issued by this laboratory are traceable to national primary standards and International System of Units (SI).

2、证书未经本机构书面授权,不得部分复制此证书。

The certificates can not be partly copied without approval of the institute.

3、本次校准结果只对此被测样品有效。

The results are only responsible for the calibrated items.

4、本次校准所使用的主要测量仪器

Main measurement standards used in this calibration

名称&编号 Name and No.	型号 / 规格 Type or size	不确定度/准确度等级/最大允许误差 Uncertainty, Accuracy or Maximum permissible error	溯源单位&证书编号 Calibration Agency and Certificate No.	有效期至 Valid until
示波器	TDS3054B	/	深圳市计量质量检测研究院 /JL2224994631	2023-08-07

5、校准地点及环境条件

Location and environmental condition for the calibration

地点: Location	本公司电学室		
温度: Temperature	22.5 °C	相对湿度: Relative Humidity	52 %
		其它: Others	/

6、本次校准所依据的技术文件(代号、名称)

Reference documents for the calibration (code、name)

参照JJG 840-2015《函数发生器检定规程》

7、敬告:

Suggestions

1) 本测量设备修理后,请立即进行校准。

Recalibrate the instrument immediately after it has been repaired.

2) 在使用过程中,如对被校准测量设备的计量特性产生怀疑,请重新校准。

Recalibrate the instrument when any suspicion about its performance arises.

8、备注(Comments):

/



校准结果

Result of Calibration

1、外观以及一般性检查: 符合要求

In view of External and Generality check: Pass

2、仪器“输出”的“+”、“-”端子间输出信号测试:

Output signal test between terminals of "+" and "-" of instrument

仪器控制/Instrument control		输出测量结果/Output measurement results		
输出调节	输出指示	频率 (Hz)	电压 (V _{p.p})	波形
Output adjustment	Output instructions	Frequency	Voltage	Waveform
最小	0	50.0	159.1	锯齿波
最大	0	50.0	158.4	锯齿波



本次校准结果的不确定度: 电压: $U_{rel}=1.3\%$, $k=2$

频率: $U_{rel}=1.3\%$, $k=2$

安正计量检测

备注 (Notes):

1.校准结论: 所校项目符合JJG 840-2015技术要求

Cal Conclusion: The calibration project is conform with the technical requirements of JJG 840-2015

2.为确保校准结果的可信度, 通常情况下建议再校周期为12个月。

To ensure the credibility of the calibration results, it is generally recommended to have a recalibration period of 12 months.

(以下空白)

(The below is blank)